

File No.: 2025-1003

January 13, 2026

s.22(1)

Dear s.22(1)

Re: **Request for Access to Records under the Freedom of Information and Protection of Privacy Act (the "Act")**

I am responding to your request of December 20, 2025 under the ***Freedom of Information and Protection of Privacy Act*** for:

Record of calibration reports for the following equipment:

1. **Sound Level Meter Model 2200 serial # 3E40214262, Sound Level Calibrator model QC-10 serial # AC300016265; and**
2. **Sound Level Meter Model 2200 serial # SAH050009, Sound Level Calibrator model QC-10 serial # QIH050069.**

Date range: January 1, 2024 to December 19, 2025.

All responsive records are attached.

Under Part 5 of the Act, you may ask the Information & Privacy Commissioner to review any matter related to the City's response to your FOI request by writing to: Office of the Information & Privacy Commissioner, info@oipc.bc.ca or by phoning 250-387-5629.

If you request a review, please provide the Commissioner's office with: 1) the request number (2025-1003); 2) a copy of this letter; 3) a copy of your original request; and 4) detailed reasons why you are seeking the review.

Yours truly,

Kevin Tuerlings, FOI Case Manager, for

[Signed by Kevin Tuerlings]

Siân Madsen, MA, MAS
Acting Director, Access to Information & Privacy

If you have any questions, please email us at foi@vancouver.ca and we will respond to you as soon as possible. You may also contact 3-1-1 (604-873-7000) if you require accommodation or do not have access to email.

Encl. (Response Package)

:kt



5613 70 St NW
 Edmonton, AB T6B 3P6
 1-888-207-2212
www.conceptcontrols.com

Certificate of Calibration

Certificate Number: WO09650SE40214262

Submitted by:	City of Vancouver 515 West 10th Avenue Vancouver, BC, V5Z 4A8	Date Issued:	19-Sep-25
Order Number:	WO09650	Valid until:	19-Sep-26
Model:	SE-402	Serial number:	SE40214262
Sub-Assemblies:	QE7052 Microphone Type 2 Preamp	Serial number:	53769
		Serial number:	n/a

Test Condition:		Model condition:	
Temperature :	18 C to 29 °C	As Found:	Out of Tolerance
Humidity:	20% to 80%	As Left:	In Tolerance
Barometric Pressure:	890mbar to 1050mbar		

Calibration per Procedure : V073-706

Reference Standard(s):		Last Calibration:	Calibration Due:
ID Number:	Device:		
CCI0001	B & K Ensemble	16-Jun-25	16-Jun-26
CCI0004	Quest Cal	10-Dec-24	10-Dec-25
CCI0015	DS360 Function Generator	17-Mar-25	17-Mar-26
CCI0014	8845A Multimeter	25-Nov-24	25-Nov-25

Measurement Uncertainty:
 $\pm 1.1\%$ Accoustic (0.1DB) $\pm 1.4\%$ VAC $\pm 0.12\%$ Hz
 Estimated at 95% confidence level (k=2)

Calibrated By: 

 Brittany Underhill 19-Sep-25

This report certifies that all calibration equipment used in the test is traceable to NIST, and applies only to the unit identified under equipment above. This report must not be reproduced except in its entirety without the written approval of Concept Controls Inc.



5613 70 St NW
 Edmonton, AB T6B 3P6
 1-888-207-2212
www.conceptcontrols.com

Certificate of Calibration

Certificate Number: WO09650AC300016265

Submitted by:	City of Vancouver 515 West 10th Avenue Vancouver, BC, V5Z 4A8	Date Issued:	19-Sep-25
		Valid until:	19-Sep-26
Order Number:	WO09650		
Model:	AC-300	Serial number:	AC300016265
Sub-Assemblies:		Serial number:	n/a
		Serial number:	n/a

Test Condition:		Model condition:	
Temperature :	18 C to 29 °C	As Found:	Out of Tolerance
Humidity:	20% to 80%	As Left:	In Tolerance
Barometric Pressure:	890mbar to 1050mbar		

Calibration per Procedure : TSI/Quest Audiometric Measurement Ensemble - Rev D

Reference Standard(s):			
ID Number:	Device:	Last Calibration:	Calibration Due:
CCI0001	B & K Ensemble	16-Jun-25	16-Jun-26

Measurement Uncertainty:
 $\pm 1.1\%$ Accoustic (0.1DB) $\pm 1.4\%$ VAC $\pm 0.12\%$ Hz
 Estimated at 95% confidence level (k=2)

Calibrated By: 

 Brittany Underhill 19-Sep-25

This report certifies that all calibration equipment used in the test is traceable to NIST, and applies only to the unit identified under equipment above. This report must not be reproduced except in its entirety without the written approval of Concept Controls Inc.